


PAT solutions for pharmaceutical industry

Alexey L. Pomerantsev



*Semenov Institute
of
Chemical Physics*



*Russian
chemometric
society*

Content

- Introduction
- Case study 1: Incoming Inspection
- Case study 2: Process Control
- Case study 3: Outcoming Inspection

Process Analytical Technology

PAT is a system for monitoring and controlling manufacturing processes (including processing) of raw and in-process materials, with the goal of ensuring final product quality.

Guidance for Industry
PAT — A Framework for
Innovative Pharmaceutical
Development, Manufacturing,
and Quality Assurance
Pharmaceutical CGMPs

Guidance for Industry PAT — A Framework for Innovative Pharmaceutical Development, Manufacturing, and Quality Assurance


U.S. Department of Health and Human Services
Food and Drug Administration
Center for Drug Evaluation and Research (CDER)
Center for Veterinary Medicine (CVM)
Office of Regulatory Affairs (ORA)

Pharmaceutical CGMPs
September 2004





and controlling
its (i.e., during
ce attributes of
s, with the goal

Pharmaceutical

PAT Instruments



Near Infra Red spectroscopy



Generic PAT Solutions

Substance

Process

Product



**Incoming
inspection**



**Process
control**



**Outcoming
inspection**

Case study 1: Incoming Inspection




To design a quick and simple routine procedure for recognition the perfect raw pharmaceutical substances directly in a warehouse

The goal


Material: Taurine, a non-essential sulfur-containing amino acid

Data: NIR spectra measured in 4100 -10000 cm^{-1} region, resolution 2 cm^{-1}


Substance: in the closed PE bags, 82 drums measured 3 times, totally: 246 spectra



Spectra



Explorative analysis




The problem:


60 points of 246 are outliers.

Are they fakes?


Reason of failure



Probe position effect




Two PCA models



Model 1


Dots are the training samples (Group 1)
Squares are the test samples (Group 2)



Model 2

Squares are the training samples (Group 1)
Dots are the test samples (Group 2)

Routine testing procedure



Case study conclusions

- Probing robust method
- Qualitative trihotomy analysis: yes / no / try again
- 100 % inspection (≤ 3 trials)

ANALYTICA CHIMICA ACTA 642 (2009) 222–227



ELSEVIER

available at www.sciencedirect.com



journal homepage: www.elsevier.com/locate/aca



Quality control of packed raw materials in pharmaceutical industry

O.Ye. Rodionova^{a,*}, Ya.V. Sokovikov^b, A.L. Pomerantsev^a

^a Semenov Institute of Chemical Physics, Kosygin 4, 119991 Moscow, Russia

^b SchelTec AG, Kosygin 19, 119334 Moscow, Russia

Case study 2: Process Control




To predict the drug release profiles during a running pellet coating process from the in-line near infrared (NIR) measurements




Objects: Pellets

Sugar +API



Coating: Acryl EZE

Fluid bed coating





Experiment



NIR Spectra

Dissolution Profiles

$t = 105$



Our goal




Prediction

Data overview




		Samples								
		1	2	3	4	5	6	7	8	9
Batches	W1	25	44	62	82	99	117	136	154	171
	W2	22	37	52	67	81	98	110	124	142
	W3	18	30	41	52	62	73	85	97	114
	W4	19	36	52	67	83	98	114	129	137
	W5	18	31	42	51	61	71	79	89	105
	W6	39	70	98	127	156	188	215	246	260
	W7	19	34	48	64	79	95	111	125	140
	Y1	21	40	59	77	96	115	133	152	168
	Y2	20	30	43	55	67	82	92	105	121
	Y3	24	46	70	89	111	133	155	176	191
	Y4	26	50	74	98	122	150	171	194	209
	Y5	18	31	42	52	63	73	83	94	110
	Y6	19	34	49	64	79	94	109	124	140

Process time, min


Conventional approach



PLS2 results



Kinetic approach




Features extraction

		k_1	k_2	...	k_p
Samples	1	3.56	0.10	...	0.33
	2	4.03	0.09	...	0.66
	3	4.99	0.10	...	0.98
	4	6.13	0.15	...	1.28
	5	6.84	0.20	...	1.62
	6	7.81	0.25	...	1.93
	7	8.66	0.30	...	2.23
	8	9.49	0.33	...	2.54
	9	9.82	0.35	...	2.54

Y

Autocatalysis


$$\varphi(t, m, k) = 100k \frac{\exp[(m+k)t] - 1}{m+k \exp[(m+k)t]}$$



$$[A] + [B] = 100$$

$$[B](0) = 0$$

$$\varphi = [B]$$



Fitter software

	A	B	C	D	E	F	G	H	I	L	M	N	O	P	Q	R	S	T	U	V	W	X
1																						
2																						
3																						
4																						
5																						
6																						
7																						
8																						
9																						
10																						
11																						
12																						
13																						
14																						
15																						
16																						
17																						
18																						
19																						
20																						
21																						
22																						
23																						
24																						
25																						
26																						
27																						
28																						
29																						
30																						
31																						
32																						
33																						
34																						
35																						
36																						
37																						
38																						
39																						



Data							Parameters		Prior Information													
x	t	Run	M	y	w	f	k1		Name	Value	Matrix											
1	0	1	0.377			0	4E-04		k1	3E-04	5E+09	0	0	0	0	0	0	0	0	0	0	0
1	15	1	0.377	97.64	1	97.58	2.402			0	0	0	0	0	0	0	0	0	0	0	0	0
1	30	1	0.377	98.58	1	99.95	3.447			0	0	0	0	0	0	0	0	0	0	0	0	0
1	45	1	0.377	96.34	1	100	4.439			0	0	0	0	0	0	0	0	0	0	0	0	0
1	60	1	0.377	100	1	100	5.221			0	0	0	0	0	0	0	0	0	0	0	0	0
1	90	1	0.377	86.87	0	100	5.774			0	0	0	0	0	0	0	0	0	0	0	0	0


```


autocatalysis
'D[y]/D[t]=(100-y)*(k1*y+k2); y(0)=0
y=k2/k1*[exp(K*t)-1]/[1+k2/k1/A0*exp(K*t)]
K=k1*(A0+k2/k1)
A0=100
k2=exp(-Lk2)
Lk2=Lk21*imp(x-1)+Lk22*imp(x-2)+Lk23*imp(x-3)+Lk24*imp(x-4)+Lk25*imp(x-5)+Lk26*imp(x-6)
k1=?
Lk21=?
Lk22=?
Lk23=?
Lk24=?
Lk25=?
Lk26=?
    
```


	Final	Deviati
1	4E-04	4E-05
2	2.402	0.247
3	3.447	0.166
4	4.439	0.165
5	5.221	0.192
5	5.774	0.226




Parameter m is common within a batch



Parameter m and the layer grade



Parameter k and the layer thickness





Intermediate conclusions

parameter m
reflects the material grade



parameter k
depends on the layer thickness

parameter k
keeps track of batch variations



Prediction of k : NLR – NIR (White subset)



Prediction of k : NLR – NIR (Yellow subset)




Test set validation: W2 and Y5 prediction



Conclusions

- PAT solution for the in-line release profile prediction
- novel “curve to curve” calibration approach via NLR
- autocatalytic model for the drug release

Details



IN-LINE PREDICTION OF DRUG RELEASE PROFILE FOR pH-SENSITIVE COATED PELLETS

Alexey L. Pomerantsev^{1,2}, Oxana Ye. Rodionova¹, Michael Melichar³,
Anthony J. Wigmore³ and Andrey Bogomolov⁴

Case study 3: Outcoming Inspection



To design a quick and simple routine procedure for the in-line inspection of the finished pharmaceutical products

Two years ago: April 2009

Mildronate



**Soft remedy for
treatment of the heart
and blood vessel
diseases**




Listenon



**Strong drug applied
for the muscle
relaxation in
anesthesia and for
intensive care**

4% aqueous solution of dexamethasone (glucocorticosteroid remedy)



Genuine objects


batch **G1**: 15 ampoules

batch **G2**: 15 ampoules

Counterfeit objects

batch **F2**: 15 ampoules

NIR layout



Bomem 160 FT NIR


spectral range 5500 - 10000 cm^{-1}

resolution 8 cm^{-1}


8 mm vial holder

T= 30°C


NIR spectra




PCA analysis



SIMCA classification




Training set is G1




Training set is G1+G2

Wet chemistry analysis




HPLC- DAD Chromatograms $\lambda=254$ nm



Peak areas of the common impurities

genuine sample G1 is used as the reference
(HPLC-DAD-UV at $\lambda=254$ nm)



Case study conclusions

- We do not check composition, e.g. the API concentration
- Qualitative analysis: yes / no
- 100 % inspection

Details

Anal Bioanal Chem (2010) 397:1927–1935

DOI 10.1007/s00216-010-3711-y

ORIGINAL PAPER

Noninvasive detection of counterfeited ampoules of dexamethasone using NIR with confirmation by HPLC-DAD-MS and CE-UV methods

**Oxana Rodionova • Alexey Pomerantsev •
Lars Houmøller • Alexey Shpak • Oleg Shpigun**

Collaborators



**Oxana
Rodionova**
ICP RAS
Russia



**Lars
Houmøller**
Arla Foods
Denmark



**Oleg
Shpigun**
MSU
Russia



**Andrey
Bogomolov**
J&M
Germany

Thanks!

